



# UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE  
United States Patent and Trademark Office  
Address: COMMISSIONER FOR PATENTS  
P.O. Box 1450  
Alexandria, Virginia 22313-1450  
www.uspto.gov

\$

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
09/770,603	01/26/2001	Arthur E. Schwartz	20076-53	8546

7590 11/17/2004

WILLIAM H. DIPPert  
REED SMITH L.L.P.  
599 LEXINGTON AVENUE  
29th FLOOR  
NEW YORK, NY 10022

EXAMINER
----------

ROLLINS, ROSILAND STACIE

ART UNIT	PAPER NUMBER
----------	--------------

3739

DATE MAILED: 11/17/2004

Please find below and/or attached an Office communication concerning this application or proceeding.

## Office Action Summary

Application No.

09/770,603

Applicant(s)

SCHWARTZ, ARTHUR E.

Examiner

Rosiland S Rollins

Art Unit

3739

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --

### Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If the period for reply specified above is less than thirty (30) days, a reply within the statutory minimum of thirty (30) days will be considered timely.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

### Status

- 1) ☒ Responsive to communication(s) filed on 8/13/04.
- 2a) ☒ This action is **FINAL**. 2b) ☐ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

### Disposition of Claims

- 4) ☒ Claim(s) 20-32 and 47-53 is/are pending in the application.
- 4a) Of the above claim(s) \_\_\_\_\_ is/are withdrawn from consideration.
- 5) ☐ Claim(s) \_\_\_\_\_ is/are allowed.
- 6) ☒ Claim(s) 20-32 and 47-53 is/are rejected.
- 7) ☐ Claim(s) \_\_\_\_\_ is/are objected to.
- 8) ☐ Claim(s) \_\_\_\_\_ are subject to restriction and/or election requirement.

### Application Papers

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☐ The drawing(s) filed on \_\_\_\_\_ is/are: a) ☐ accepted or b) ☐ objected to by the Examiner.  
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).  
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

### Priority under 35 U.S.C. § 119

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☐ All b) ☐ Some \* c) ☐ None of:
- ☐ Certified copies of the priority documents have been received.
  - ☐ Certified copies of the priority documents have been received in Application No. \_\_\_\_\_.
  - ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

\* See the attached detailed Office action for a list of the certified copies not received.

### Attachment(s)

- |  |   |
|--|---|
| 1) <input type="checkbox"/> Notice of References Cited (PTO-892)   | 4) <input type="checkbox"/> Interview Summary (PTO-413)<br>Paper No(s)/Mail Date. _____ |
| 2) <input type="checkbox"/> Notice of Draftsperson's Patent Drawing Review (PTO-948)                                   | 5) <input type="checkbox"/> Notice of Informal Patent Application (PTO-152)             |
| 3) <input type="checkbox"/> Information Disclosure Statement(s) (PTO-1449 or PTO/SB/08)<br>Paper No(s)/Mail Date _____ | 6) <input type="checkbox"/> Other: _____  |

## **DETAILED ACTION**

### ***Claim Rejections - 35 USC § 102***

The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this Office action:

A person shall be entitled to a patent unless –

(b) the invention was patented or described in a printed publication in this or a foreign country or in public use or on sale in this country, more than one year prior to the date of application for patent in the United States.

Claims 20-28, 30, 31, 47-51 and 53 continue to be rejected under 35 U.S.C. 102(b) as being anticipated by Saab (5,624,392). Saab shows a catheter (10) for intravascular corporeal cooling comprising: an elongated tubular member (12) having proximal and distal sections, an outer surface, and at least one lumen (11) extending therethrough, and annular insulation (16, 22) having proximal and distal ends arranged concentrically (Fig. 1) around the outer surface of the elongated tubular member (12) is insulated from fluid or tissue external to the annular insulation (16, 22), wherein the insulation is tapered (at 18), extends along substantially the whole length or a shorter section of the catheter (col 16, lines 54-59) and comprises a fluid-filled member, filled with gas, water or saline (col 9, line 66 - col 10 line 2) and polymeric material (14, 20, col 10, lines 38-39). Regarding claim 30, the structure can be used for brain cooling.

### ***Claim Rejections - 35 USC § 103***

The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and

Art Unit: 3739

the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negated by the manner in which the invention was made.

Claims 29 and 52 are rejected under 35 U.S.C. 103(a) as being unpatentable over Ginsburg in view of Donlon (6,056,723). Ginsburg shows all of the limitations of claim 29 except for the pressure sensor. Donlon shows a similar catheter for placement in a blood vessel, which includes a pressure sensor (38). It would have been obvious to one of ordinary skill in the art at the time of the invention to use a pressure sensor in the device of Ginsburg to prevent injury by which can be caused by high pressure fluids in the bloodstream.

### ***Response to Arguments***

1. Applicant's arguments filed 8/13/04 have been fully considered but they are not persuasive.

Applicant argues that Saab allows the temperature of the fluid inside the catheter to be transferred to adjacent portions of the body that are in contact with or in proximity to the catheter sidewalls. In col. 11 lines 32-34 Saab describes an insulation feature in which the portions of the body that are in contact with the catheter sidewalls are insulated from the cool fluid flowing through the catheter.

In response to applicant's argument that the references fail to show certain features of applicant's invention, it is noted that the features upon which applicant relies (i.e., a system where any portion of the fluid could remain in the body) are not recited in the rejected claim(s). Although the claims are interpreted in light of the specification,

Art Unit: 3739

limitations from the specification are not read into the claims. See *In re Van Geuns*, 988 F.2d 1181, 26 USPQ2d 1057 (Fed. Cir. 1993).

Regarding the argument that Donlon does not show or teach a pressure sensor at or adjacent to the distal end of the catheter, in figure 2 Donlon illustrates a pressure port (37), which is located at the distal end of the catheter.

**THIS ACTION IS MADE FINAL.** Applicant is reminded of the extension of time policy as set forth in 37 CFR 1.136(a).

A shortened statutory period for reply to this final action is set to expire **THREE MONTHS** from the mailing date of this action. In the event a first reply is filed within **TWO MONTHS** of the mailing date of this final action and the advisory action is not mailed until after the end of the **THREE-MONTH** shortened statutory period, then the shortened statutory period will expire on the date the advisory action is mailed, and any extension fee pursuant to 37 CFR 1.136(a) will be calculated from the mailing date of the advisory action. In no event, however, will the statutory period for reply expire later than **SIX MONTHS** from the mailing date of this final action.

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Rosiland S Rollins whose telephone number is 703/3082711. The examiner can normally be reached on Mon.-Fri. 9:00 AM - 6:00 PM.

Art Unit: 3739

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Linda C. Dvorak can be reached on 703/3080994. The fax phone number for the organization where this application or proceeding is assigned is 703-872-9306.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free).

  
Rosiland S Rollins  
Primary Examiner  
Art Unit 3739

RR